



Equipment Service Scope of Works & Spare Parts

ATI Q45H Chlorine & pH Analyser	Major Service Yearly (mandatory)	Minor Service Twice Yearly (recommended)
▪ Replace Membrane & Electrolyte in Chlorine Probe	✓	✓
▪ Replace Electrolyte Chamber O-rings		✓
▪ Clean Flow Cell (note any defects)	✓	✓
▪ Replace Cell Block O-ring		✓
▪ Calibrate Analyser	✓	✓
▪ Check Operation, Functions & Alarms	✓	✓
▪ Check Operation of Flow Switch & Sample Water Flow Regulator	✓	✓
▪ Check & Clean Super Y Filter	✓	✓

Major Service Parts:

Membrane & Electrolyte

Minor Service Parts:

Membrane & Electrolyte

Membrane and electrolyte should be changed to suit water quality and will vary from site to site

WESTERN AUSTRALIA

9 Chifley Brace
Jandakot WA 6164

Telephone: (08) 9412 6100

Fax: (08) 9412 6199

Email: hydramet@hydramet.com.au

SOUTH AUSTRALIA

29 Deloraine Road
Edwardstown SA 5039

Telephone: (08) 8374 7800

Fax: (08) 8374 7899

Email: hydrasa@hydramet.com.au

VICTORIA

Unit 4, 47-51 Little Boundary Road
Laverton North VIC 3026

Telephone: (03) 9325 3900

Fax: (03) 9318 9653

Email: hydravic@hydramet.com.au